

UCD Genome Center	SOP Standard Operating Procedure [Under Development]	Page 1 of 2
date: 05/28/2013	Extraction of Mammalian Tissue Samples: Lungs/Muscle/Heart	Code No.: tissue 05282008

Issued: 05/28/2008 Valid From: Not yet valid	Validity Area: UC Davis Genome Center, Metabolomics Core and Research Laboratories
Responsible:	
This SOP supersedes: SOP 03-04-2008	Approved:

Extraction of Mammalian Tissue Samples: Lungs/Muscle/Heart

1. References:

- Fiehn O, Kind T (2006) Metabolite profiling in blood plasma. In: Metabolomics: Methods and Protocols. Weckwerth W (ed.), Humana Press, Totowa NJ (in press)

2.Starting material:

- Mammalian tissue: Lung/Muscle/Heart: Whole tissue sample is prepared OR stein mill whole tissue sample and weigh 50mL aliquot.

3. Equipment:

- Centrifuge (Eppendorf 5415 D)
- Calibrated pipettes 1-200µl and 100-1000µl
- Eppendorf tubes 2ml, uncoloured (Cat. No. 022363204)
- Centrifuge tubes, various sizes, polypropylene
- Eppendorff Tabletop Centrifuge (Proteomics core Lab.)
- ThermoElectron Neslab RTE 740 cooling bath at -20°C
- MiniVortexer (VWR)
- Orbital Mixing Chilling/Heating Plate (Torrey Pines Scientific Instruments)
- Speed vacuum concentration system (Labconco Centrivap cold trap)
- Turex mini homogenizer

4. Chemicals

- Acetonitrile, LCMS grade (JT Baker; Cat. No.9829-02)
- Isopropanol, HPLC grade (JT Baker; Cat. No. 9095-02)
- Crushed ice
- pH paper 5-10 (EMD Chem. Inc.)
- Nitrogen line with pipette tip
- 18 MΩ pure water (Millipore)

5. Procedure

Preparation of extraction mix and material before experiment:

- Switch on bath to pre-cool at -20°C (±2°C validity temperature range)
- Check pH of acetonitrile and isopropanol (pH7) using wetted pH paper
- Make the extraction solution by missing acetonitrile, isopropanol and water in proportions 3 : 3 : 2

UCD Genome Center	SOP Standard Operating Procedure [Under Development]	Page 2 of 2
date: 05/28/2013	Extraction of Mammalian Tissue Samples: Lungs/Muscle/Heart	Code No.: tissue 05282008

4. Rinse the extraction solution for 5 min with nitrogen. Make sure that the nitrogen line was flushed out of air before using it for degassing the extraction solvent solution

Sample Preparation

1. Weigh 4 mg tissue sample (muscle 20 mg) in to a 2.0 ml eppendorf tube. Add 1.0 mL extraction solvent to the tissue sample and using GenoGrinder homogenize samples for 45 seconds ensuring that sample resembles a powder.
2. Centrifuge the samples at 2500 rpm. for 5 minutes. Aliquot 2 X 500µl supernatant, one for analysis and one for a backup sample. Store backup aliquot in the -20°C freezer.
3. Evaporate one 500µl aliquot of the sample in the Labconco Centrivap cold trap concentrator to complete dryness
4. The dried aliquot is then re-suspended with 500µl 50% acetonitrile (degassed as given) (only for liver and brain samples).
5. Centrifuge for 2 min at 14000 rcf using the centrifuge Eppendorf 5415.
6. Remove supernatant to a new Eppendorff tube.
7. Evaporate the supernatant to dryness in the the Labconco Centrivap cold trap concentrator.
8. Submit to derivatization.

The residue should contain membrane lipids because these are supposedly not soluble enough to be found in the 50% acetonitrile solution. Therefore, this 'membrane residue' is now taken for membrane lipidomic fingerprinting using the nanomate LTQ ion trap mass spectrometer. Likely, a good solvent to redissolve the membrane lipids is e.g. 75% isopropanol (degassed as given above). If the 'analysis' aliquot is to be used for semi lipophilic compounds such as tyrosine pathway intermediates (incl. dopamine, serotonin etc, i.e. polar aromatic compounds), then these are supposedly to be found together with the 'GCTOF' aliquot. We can assume that this mixture is still too complex for Agilent chipLCMS. Therefore, in order to develop and validate target analysis for such aromatic compounds, we should use some sort of Solid Phase purification. We re-suspend the dried 'GCTOF' aliquot in 300 µl water (degassed as before) to take out sugars, aliphatic amino acids, hydroxyl acids and similar logP compounds. The residue should contain our target aromatics. We could also try to adjust pH by using low concentration acetate or phosphate buffer. The residue could then be taken up in 50% acetonitrile and used for GCTOF and Agilent chipMS experiments. The other aliquot should be checked how much of our target compounds would actually be found in the 'sugar' fraction.

6. Problems

To prevent contamination disposable material is used. Control pH from extraction mix.

7. Quality assurance

For each sequence of sample extractions, perform one blank negative control extraction by applying the total procedure (i.e. all materials and plastic ware) without biological sample.

8. Disposal of waste

Collect all chemicals in appropriate bottles and follow the disposal rules.